

EQUIPMENT PURCHASE SPECIFICATION

FOR BME REPRESENTATIVE USE ONLY	
Approved By	Name : Zaenul Rosli
	Signature :
Date	16/7/2026

TITLE OF TENDER/QUOTATION	TO SUPPLY, DELIVER, TEST AND COMMISSION TWO (2) UNITS OF MULTI DISPENSER PIPETTE FOR FORENSIC BIOLOGY LABORATORY , DEPARTMENT OF SCIENTIFIC SERVICES, MINISTRY OF HEALTH.
----------------------------------	--

TENDER REFERENCE NO.	
-----------------------------	--

NAME OF ITEM	MULTI DISPENSER PIPETTE
---------------------	-------------------------

NO.	ITEM DESCRIPTIONS AND SPECIFICATIONS	VENDOR'S OFFER			
		COMPLY (Please tick v) (Provide evidence for compliance(s))		ITEM DESCRIPTIONS AND SPECIFICATIONS	PRICE (\$)
		YES	NO		
1.0	<u>GENERAL</u>				
1.1	Tenderer is to supply and provide the following: - Two (2) units of multi dispenser pipettes - Warranty, preventive maintenance and calibration				
2.0	<u>MULTI DISPENSER PIPETTE</u>				
2.1	Number of channels : 1-Channel				
2.2	Pipetting Type: Positive displacement system				
2.3	High speed dispensing with motorised piston				

Procurement Ref. No. : DSS/FS/PROC/B50K/2026-2027/032(FOR)
Section/Laboratory Ref. No. : DSS/FOR/PR/2026/14

2.4	Automatic pipette tips recognition				
2.5	One button tip ejector: for one-handed operation and contact-free tips advance ejection				
2.6	Volume range from 1 µL to 50 ml				
2.7	Selectable dispensing speed				
2.8	Operating mode: electronic				
2.9	Illuminated display				
2.10	Rechargeable lithium-ion battery				
2.11	Charger stand with charging Cable. All electrical and electronic components in the system must be able to utilise mains power supply of 220 – 240 V AC, 50 – 60 Hz. Otherwise, tenderer to include additional items or works required to ensure successful operation.				
3.0	<u>WARRANTY, PREVENTIVE MAINTENANCE AND CALIBRATION</u>				
3.1	A minimum of one (1) year warranty for manufacturer's defect on the hardware, software and all cost of repairs and/or replacements should be included.				
3.2	After-sales services must be provided for the product after one (1) year.				
3.3	One-off preventive maintenance to be carried out just before or soon after the one-year warranty period. Scope of work to follow manufacturer's				

	manual / recommendation specific for the equipment offered, which include: • Supply, delivery and installation of preventive maintenance kits and/or consumables • Software update (to obtain prior authorization from user and BME) • Inspection • cleaning • alignment • any other related preventive maintenance works required				
3.4	Pipette to be calibrated by ISO 17025:2017 accredited Calibration Service Centre prior to delivery and one (1) year after delivery. Calibration will be done using Combitips Advanced 1 mL and 2.5 mL. Parameters for the calibration includes calibration points as below: • 10 µL • 15 µL • 18 µL • 25 µL • 10% of nominal (100 µL or 250 µL) • 50% of nominal (500 µL or 1250 µL) • Nominal (1 mL or 2.5 mL)				
4.0	<u>DELIVERY</u> Items offered MUST be delivered within _____ from date of approval. (Vendor to indicate the delivery period.)				
5.0	<u>PRICE VALIDITY</u> Price validity MUST NOT BE LESS THAN six (6) months.				

	TOTAL PRICE (\$)
--	-------------------------

No.	GENERAL SPECIFICATIONS*	VENDOR'S OFFER
A	Total Price: B\$	
B	Delivery Period:	
C	Model & Brand:	
D	Country of Origin:	
E	Where marketed:	
F	Year of Manufacture:	
G	Warranty:	
H	Power Requirement:	
I	Battery Back-up:	
J	International Safety Standard:	
K	Technical Support:	
L	Equipment Whole Life Support:	

Procurement Ref. No. : DSS/FS/PROC/B50K/2026-2027/032(FOR)
Section/Laboratory Ref. No. : DSS/FOR/PR/2026/14

No.	GENERAL SPECIFICATIONS*	VENDOR'S OFFER
M	Dimensions (WxHxD) cm:	
N	User Manuals:	
O	Service Manuals:	
P	Spare-parts & Consumables Listing:	
Q	Technical Training On-Site:	
R	Site Requirements:	

*To all participating companies, please fill in the table above along with your other documents during submission of tender.

Bahagian/Unit/Makmal :	Makmal Forensik Biologi		
Bil. Rujukan Bahagian/Unit/Makmal :	DSS/FOR/PR/2026/14		
Pegawai dirujuk :	Nama :	Nuramalina Sawal	
	E-mail :	Nuramalina.sawal@moh.gov.bn	
	Tel.No.:	2382424 ext 6068	Fax No. : -