

UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING,
 TWENTY-FIVE (25) UNITS OF SMALL WATER DISTILLER FOR VARIOUS DENTAL CLINIC,
 DEPARTMENT OF DENTAL SERVICES, MINISTRY OF HEALTH**

NO.	TERMS AND CONDITIONS	VENDOR'S OFFER (PLEASE STATE)
1	Tenderer must be registered with the Ministry of Health.	
2	TENDER FORM should be filled completely including the USER REQUIREMENT FORM (if available). Submission of incomplete form MAY cause DISQUALIFICATION OF TENDER.	
3	Each tenderer is allowed to quote ONE BRAND WITH ONE PRICE ONLY for each item. Submission of more than one brand and price will cause DISQUALIFICATION OF TENDER.	
4	All consumables supplied throughout this tender shall have a minimum expiry date of twelve (12) months / on delivery. Should the consumables be urgently needed, provision of consumables with expiry date of less than twelve (12) months should be first agreed by the User before delivery is made.	
5	Brochures / catalogues should be submitted / attached with tender document.	
6	Any room renovation which may be required, it is mandatory to conduct site visit (if applicable)	
7	DELIVERY PERIOD: (Please state) Not More Than 90 days upon confirmation	(Yes / No) (If No, please specify)
8	PRICE VALIDITY: The quotation shall remain valid for 12 MONTHS from the final date for the submission of the quotation and no supplier may withdraw his/her quotation within that period. The Government reserves the right to extend this period if deemed necessary provided that such extension to the quotation validity period shall have written consent of the supplier(s).	

UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING,
 TWENTY-FIVE (25) UNITS OF SMALL WATER DISTILLER FOR VARIOUS DENTAL CLINIC,
 DEPARTMENT OF DENTAL SERVICES, MINISTRY OF HEALTH**

Distribution list	1. National Dental Centre, Block A (5 units) 2. Berakas Health Centre (3 units) 3. Suri Seri Begawan Hospital, Kuala Belait (2 units) 4. PIHM Hospital, Temburong (1 unit) 5. PMMPMHAB Hospital, Tutong (2 units) 6. PAPHRs Sungai Asam Health Centre (2 units) 7. School Dental Services (4 units) 8. National Dental Centre, Block C (2 units) 9. Primary Oral Care Services (4 units)
--------------------------	--

SECTION I - USER REQUIREMENT		YES	NO	PLEASE SPECIFY THE SPECIFICATION/ SCOPE OFFERED FAILURE TO FILL THE OFFER DETAIL SHALL BE IDENTIFIED AS NON-COMPLIANCE TO THE OFFER
Small Water Distiller				
Standard features	Quantity: Twenty-five (25) units			
	Portable			
	To produce distilled water via evaporation and steam condensation process in a controlled environment			
	Capacity: 4 litres			
	Water production: At least 1 litre/hour			
	Total dissolved solids: Must consistently output 0 to 1ppm			
	ISO 13485 (Medical Devices)			
	Chamber material: 316 Marine Grade Stainless Steel			
	Collection Container (Carafe): High-durability Medical Polypropylene (PP) or Borosilicate Glass			
	Automatically switched off once distilling process is completed.			
	Features an automatic overheat shut-off that kills power the exact moment the chamber runs dry.			
	Cooling Fan: Aluminium fan blades.			
	Easy to clean			
	Accessories to be included (The Equipment Should Come With All Necessary Accessories For It To Be Fully Operational)			

UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING,
 TWENTY-FIVE (25) UNITS OF SMALL WATER DISTILLER FOR VARIOUS DENTAL CLINIC,
 DEPARTMENT OF DENTAL SERVICES, MINISTRY OF HEALTH**

SECTION II - PROCUREMENT REQUIREMENT

Model and Brand of Equipment.	
Country of Origin.	
Unit (B\$)	
Total Price Offered (CIF)	
Country of Origin	
Year manufactured.	
Warranty period (at least 1 year)	_____ years
Delivery period (not more than 90 days)	_____ weeks
Price validity (not less than 12 months)	_____ years

UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING,
 TWENTY-FIVE (25) UNITS OF SMALL WATER DISTILLER FOR VARIOUS DENTAL CLINIC,
 DEPARTMENT OF DENTAL SERVICES, MINISTRY OF HEALTH**

TECHNICAL SPECIFICATIONS/REQUIREMENTS	
Mains Supply 240VAC 50 Hz.	
State equipment ambient operating temperature range.	
The equipment must comply to either one of the five international safety standards, namely; US FDA Standard, European Union CE MARK, Australian TGA Standard, Canadian CSA Standard or Japanese JIS Standard.	
State number of locally stationed engineers/technicians and the nearest overseas support.	
State overall dimensions of equipment.	
State net weight of the equipment (kg)	
Equipment whole life time support Supplier must guarantee spare parts are available at least for 5 years after the installation of the equipment.	_____years
Brochure Tenderers must submit with their offer, a detailed brochure which supports the information the tenderer has entered in the "form-to-be-used (FORM B)" for reference.	
User and Service Manuals Two hard copies and one soft copy [PDF FORMAT] to be submitted to BME office during commissioning.	
Spare-parts and consumables Comprehensive spare-parts and consumables listings in the form of softcopy must be provided. Listing must have the following details; Part Number and Part Description.	
Training Tenderers are required to conduct technical training to Biomedical Engineers and Technicians ● Training to be conducted locally, tenderers are required to: ● Provide training materials, test equipment, demo equipments, etc. ● Provide training to two groups of technical staffs.	

UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING,
TWENTY-FIVE (25) UNITS OF SMALL WATER DISTILLER FOR VARIOUS DENTAL CLINIC,
DEPARTMENT OF DENTAL SERVICES, MINISTRY OF HEALTH**

- | | |
|--|--|
| <ul style="list-style-type: none">● Provide 2 days (minimum) of training for each group.● Training to be conducted at BME office/ Hospital. | |
|--|--|