

LAMPIRAN - A1

RUJ: BME/A2K/(1750)/2026/HTD (B8)

19/9/2026

**TO REPAIR ONE (1) UNIT HD INTEGRA REVERSE OSMOSIS WATER TREATMENT SYSTEM AT WARD 17 & 18 ,
HOSPITAL RIPAS.**

SCOPE OF WORK

To repair one (1) unit of HD Integra Reverse Osmosis Water Treatment System at Ward 17 & 18 Ripas Hospital. Detail are as follow:

1. BME19413, Brand: HD Integra, Model: Aeros HD-60, SN: 2016-032, 2026-033, at Ward 17 & 18 W/order 2026-12566

- Scope of work to follow equipment manual/recommendation specific for this system which include the following where if necessary:

- (i) installation of items listed in spareparts list (Lampiran A2)
- (ii) alignment
- (iii) calibration
- (iv) optimisation
- (vi) any other related maintenance works required after installation of items listed in spareparts list (Lampiran A2)

- Perform functional test, verification and performance check post-maintenance.

- Disposal of any used spare parts, unless otherwise instructed.

- The quotation submitted shall remain valid for a period of not less than twelve (12) months from the date of submission.

- Work on the system is to be carried out by engineer(s) authorised in providing maintenance to the equipment.

LAMPIRAN - A2

RUJ: BME/A2K/(1750)/2026/HTD (B8)

TO REPAIR ONE (1) UNIT HD INTEGRA REVERSE OSMOSIS WATER TREATMENT SYSTEM AT WARD 17 & 18 ,
HOSPITAL RIPAS.

SPAREPART		PART NUMBER	QUANTITY
1	LCD TOUCH SCREEN	ATC-EL 0025	1