

EQUIPMENT PURCHASE SPECIFICATION

FOR BME REPRESENTATIVE USE ONLY	
Approved By	Name : Zaenul Rosli
	Signature :
Date	29/9/2025

TITLE OF TENDER/QUOTATION	TO SUPPLY, DELIVER, INSTALL, TEST AND COMMISSION ONE (1) UNIT OF WATER BATH FOR TOXICOLOGY LABORATORY, DEPARTMENT OF SCIENTIFIC SERVICES, MINISTRY OF HEALTH.
----------------------------------	---

TENDER REFERENCE NO.	DSS/FS/PROC/B50K/2026-2027/037(TOX)
-----------------------------	-------------------------------------

NAME OF ITEM	ELECTRIC THERMOSTAT WATER BATH
---------------------	--------------------------------

NO.	ITEM DESCRIPTIONS AND SPECIFICATIONS	VENDOR'S OFFER			
		COMPLY (Please tick v) (Provide evidence for compliance(s))		ITEM DESCRIPTIONS AND SPECIFICATIONS	PRICE (\$)
		YES	NO		
1.0	<u>General Requirments:</u>				
1.1	Suitable for sample incubation and reagent warming				
1.2	Uniform temperature distribution throughout the chamber				
1.3	Table top				
1.4	Drain valve for easy water discharge				
2.0	<u>Technical Specification:</u>				
2.1	Capacity: 10-20L				

Procurement Ref. No. : DSS/FS/PROC/B50K/2026-2027/037(TOX)
 Section/Laboratory Ref. No. : DSS/TOX/PR/2026/B50K/003

2.2	Temperature Control Range: RT +5~100°C				
2.3	Temperature Fluctuation Degree: ±0.5°C				
2.4	Temperature Resolution Ratio: 0.1°C				
2.5	Digital temperature with LED/LCD display of actual temperature and set temperature				
3.0	<u>Electrical Requirements</u>				
3.1	All electrical and electronic components in the system must be able to utilize mains power supply of 220-240V, 50 – 60 Hz. Otherwise, vendor to include additional items or works required to ensure successful operation.				
4.0	<u>Accessories</u>				
4.1	Two (2) units of Non-mercury digital Thermometers to monitor water bath temperatures. Includes features of minimum, maximum temperature and NIST-Traceable certificate. Temperature range is between 0° to 100°C.				
5.0	<u>WARRANTY</u>				
5.1	A minimum of one (1) year warranty for manufacturer's defect on the hardware, software and all cost of repairs and/or replacements should be included				
5.2	After-sales services and technical support shall continue to be provided by the vendor beyond the warranty period.				

5.3	One-off preventive maintenance to be carried out just before or soon after the one-year warranty period. Scope of work to follow manufacturer's manual / recommendation specific for the equipment offered, which include: • Supply, delivery and installation of preventive maintenance kits and/or consumables • Inspection • cleaning • alignment • calibration • any other related preventive maintenance works required				
6.0	<u>DELIVERY</u> Items offered MUST be delivered within 12 weeks from date of approval. (Vendor to indicate the delivery period.)				
7.0	<u>PRICE VALIDITY</u> Price validity MUST NOT BE LESS THAN 6 Months.				
<u>TOTAL PRICE (\$)</u>					

No.	GENERAL SPECIFICATIONS*	VENDOR'S OFFER
A	Total Price: B\$	
B	Delivery Period:	
C	Model & Brand:	
D	Country of Origin:	
E	Where marketed:	
F	Year of Manufacture:	
G	Warranty:	
H	Power Requirement:	
I	Battery Back-up:	
J	International Safety Standard:	
K	Technical Support:	
L	Equipment Whole Life Support:	
M	Dimensions (WxHxD) cm:	
N	User Manuals:	

Procurement Ref. No. : DSS/FS/PROC/B50K/2026-2027/037(TOX)
Section/Laboratory Ref. No. : DSS/TOX/PR/2026/B50K/003

No.	GENERAL SPECIFICATIONS*	VENDOR'S OFFER
O	Service Manuals:	
P	Spare-parts & Consumables Listing:	
Q	Technical Training On-Site:	
R	Site Requirements:	

*To all participating companies, please fill in the table above along with your other documents during submission of tender.

Bahagian/Unit/Makmal :	TOXICOLOGY LABORATORY		
Bil. Rujukan Bahagian/Unit/Makmal :	DSS/TOX/PR/2026/B50K/003		
Pegawai dirujuk :	Nama :	MOHAMMAD SALLEH BIN HAJI HUSSIN	
	E-mail :	salleh.hussin@moh.gov.bn	
	Tel.No.:	2382424	Fax No. : 2381946