

TENDER REF. NO.: KK/253/2026/UPP

**MINISTRY OF HEALTH
NEGARA BRUNEI DARUSSALAM**

**SUPPLY, DELIVERY, INSTALLATION, TESTING AND
COMMISSIONING OF COLPOSCOPY SYSTEM FOR
OBSTETRIC & GYNECOLOGY, WOMEN AND CHILDREN
CENTRE RAJA ISTERI PENGIRAN ANAK SALEHA
HOSPITAL**

TENDER FEE : \$10.00

RECEIPT NO. :

CLOSING DATE : ON TUESDAY, 17th November 2026

TIME : 2.00 PM

FOA :

**THE CHAIRMAN
MINI TENDER BOARD, TENDER BOX
GROUND FLOOR, MINISTRY OF HEALTH
COMMONWEALTH DRIVE
BANDAR SERI BEGAWAN BB3910
NEGARA BRUNEI DARUSSALAM**

[CLUSTERING]

SECTION 2

SPECIFICATION

TENDER REFERENCE NO.: KK/253/2026/UPP

INVITATION TO TENDER

SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING OF COLPOSCOPY SYSTEM FOR OBSTETRIC & GYNECOLOGY, WOMEN AND CHILDREN CENTRE RAJA ISTERI PENGIRAN ANAK SALEHA HOSPITAL

SCOPE OF WORK	
Please ☒ Tick where appropriate	
Supply of <u>one (1) set</u> of <u>Colposcopy</u> for Obstetrics and Gynaecology Department, RIPASH	

SECTION 1 – USER REQUIREMENTS	
1	EQUIPMENT SPECIFICATION
Please ☒ Tick where appropriate	
1.1	The system shall be a high-definition digital colposcopy system designed for examination, diagnosis, documentation, and minor procedures of the cervix, vagina, and vulva in gynaecology clinics and operating theatres.
1.2	The system shall be suitable for use in gynaecology clinic and minor procedure room
1.3	System configuration The system shall consist of, but not limited to: - Colposcope/imaging unit - Illumination system - Camera/imaging system - Medical-grade display monitor - Mobile stand - Image documentation software.

1.4	Configuration: Floor stand, Trolley Mounted with lockable wheels
1.5	Compact endoscopy system
1.5.1	Configuration: Trolley Mounted with lockable wheels and drawer
1.5.2	Display
1.5.2.1	Medical-grade monitor with high-definition display
1.5.2.2	Monitor display size: Minimum 18” or better
1.5.2.3	Touch-screen or physical control interface acceptable
1.5.3	Optical and imaging system
1.5.3.1	Integrated image processing module
1.5.3.2	Full HD camera head for full HD imaging resolution or higher
1.5.3.3	High-quality optical system with clear image reproduction and colour accuracy.
1.5.4	Integrated LED light source with adjustable light intensity control
1.5.5	Documentation and connectivity
1.5.5.1	Image and Video Output: Digital, with USB export capability
1.5.5.2	Complete with Interface for storing and exporting video files
1.5.5.3	Monitor supports data export via USB for clinical documentation
1.6	Telescope for colposcopy procedure
1.6.1	Autoclavable 90° telescope for colposcopy procedures with integrated illumination, or equivalent technology.
1.6.2	Working distance: Approximately 25-75cm or better
1.6.3	Magnification: 7.5x to 15x or better
1.6.4	Compatible with the compact endoscopy system offered
1.7	The system must be DICOM and PACS ready and enabled for future integration with Hospital Information System (HIS) where required.
1.8	The system shall comply with relevant international medical device safety standards (e.g IEC 60601-1 or equivalent).
1.9	Consumables and Accessories: The Tenderer shall supply the required accessories and consumables to ensure full functionality of the machine, including but not limited to the following: - Cover

	- Suitable sterilization and storage containers shall be supplied for all reusable components of the system requiring sterilization.
--	--

2. END USER AND TECHNICAL TRAINING	
Please ☒ Tick where appropriate	
2.1	Conduct user training to the all-end users by an application specialist or competent local engineer including but not limited to: • Basic user operation, user troubleshooting and user maintenance • Provide Operating manual (Hardcopy and/or Softcopy)
2.1.1	Tenderer must prepare a training attendance or proof of training done to end user during commissioning and the refresher course (6) months after commissioning.
2.2	Introductory Technical Training to Biomedical Engineers and Technicians at BME Office by competent Tenderer's Engineer/Technicians that includes but not limited to: • Troubleshooting and basic corrective maintenance • Handling and basic inspection maintenance *(Two sessions/groups if required)

3. WARRANTY	
Please ☒ Tick where appropriate	
3.1	Tenderer to include warranty period of at least TWO (2) years
3.2	Tenderers to ACKNOWLEDGE the Warranty Undertaking Form in Section 4 stating the terms of warranty provided for the equipment in the tender for the period of two years. This includes but not limited to: • Scope of Warranty • One-time Planned Preventive Maintenance Per Year during warranty • Comprehensive Corrective Maintenance of Main Unit

4. PRODUCT DEMONSTRATION (UPON REQUIREMENT)

Please ☒ Tick where appropriate

7.1

If requested by the Government, vendor shall provide a product demonstration as part of the evaluation process

The tenderer shall provide either of the following demonstration modes:

- Physical Demonstration (In-Person):

A complete working unit must be brought to the designated demonstration site as specified by the procuring entity. The demonstration shall be conducted by qualified personnel from the supplier.

- Online Demonstration (Virtual):

The demonstration may be conducted via a live video conferencing platform (e.g., Zoom, MS Teams). The session must include:

- Live operation of the actual product.
- Real-time interaction to address questions or perform requested functions.
- High-definition video and clear audio for full visibility and understanding.

SECTION 2 – PRICE PROPOSAL	
PURCHASE PRICE	PER UNIT
	TOTAL

SECTION 3 - PROCUREMENT AND TECHNICAL SPECIFICATION	
BRAND	MODEL
COUNTRY OF ORIGIN	YEAR INTRODUCED TO MARKET
WARRANTY PERIOD	LAST COUNTRY SOLD TO
PRICE VALIDITY [AT LEAST <u>ONE (1) YEAR</u> PRICE VALIDTY]	DELIVERY TIME

SECTION 3 - PROCUREMENT AND TECHNICAL SPECIFICATION
AUTHORIZED DISTRIBUTOR: (AUTHORIZED DISTRIBUTOR LETTER ATTACHED)
DETAILED BROCHURE INCLUDED
USER AND SERVICE MANUALS:
MAINS POWER SUPPLY:
BATTERY
POWER ADAPTER/CHARGER OUTPUT RATING:
EQUIPMENT AMBIENT OPERATING TEMPERATURE RANGE:
INTERNATIONAL SAFETY STANDARD Must comply to at least 1 safety Standards and certification (Please attached the copy of stated standards and certifications)
NUMBER OF TECHNICAL SUPPORT (ENGINEER/TECHNICIAN) Please provide training or certification for locals who is trained/certified
DIMENSIONS AND WEIGHT OF MAIN UNIT:
EQUIPMENT WHOLE LIFE TIME SUPPORT:

SECTION 4 – WARRANTY UNDERTAKING FORM

Tenderer, on behalf of the manufacturer, acknowledged and agrees that when equipment is under the warranty period, must cover the scope of normal warranty below at no additional cost:

NORMAL WARRANTY

- Warrants the supplied medical equipment and its accessories to be in good condition, in working order and free from defects to the extent such equipment do not comply with specifications, under normal use for the warranty period. The scope of warranty covers to its maximum extent permitted by applicable law.
- During warranty, tenderer must rectify issues arise from any mechanical, technical or software faulty as soon as it is reported.
- **Exchange warranty;** Providing replacement units:
 - A. Warranty against defects – Manufacturing defects or Equipment malfunction resulted from mechanical, electrical or software failure during Commissioning or within the first _____ months of use
 - B. Faulty workmanship or unsatisfactory condition during delivery or commissioning
 - C. If a unit or accessory is deemed used item or refurbished item (not a new unit) by the user and BME Unit.
- **Two time Planned Preventive Maintenance (PPM) PER YEAR** according to Manufacturer's Preventive Maintenance Guideline and to include one-time replacements of battery at the end of 2 years warranty period or any other relevant parts to prolong equipment lifespan.

EXCLUSION FROM WARRANTY

MOH understand that the following circumstances are not covered in the warranty and Tenderer may quote for repair and subject to MOH approval:

- Unauthorized modifications - an alteration or repair by anyone other than the Manufacturer or Authorized agent during warranty period.
- Accidental damage or problems caused by negligence or mishandling, subject to appropriate justification by both parties.
- Vandalism and Natural disasters
- Normal wear and tear

ANY OTHER EXCLUSION

Tenderer may propose below to include items or terms which is not listed in the exclusion list above for MOH consideration.

TENDERER ACKNOWLEDGMENT

COMPANY CHOP AND SIGNATURE

	TERMS AND CONDITIONS
1	Tenderer must be registered with the Ministry of Health.
2	TENDER FORM should be filled completely including the USER REQUIREMENT FORM (if available). Submission of incomplete form MAY cause DISQUALIFICATION OF TENDER .
3	Each tenderer is allowed to quote ONE BRAND WITH ONE PRICE ONLY for each item. Submission of more than one brand and price will cause DISQUALIFICATION OF TENDER .
4	All consumables supplied throughout this tender <u>shall</u> have a minimum expiry date of twelve (12) months / on delivery (if applicable). Should the consumables be urgently needed, provision of consumables with expiry date of less than twelve (12) months should be first agreed by the User before delivery is made (if applicable).
5	Brochures / catalogues should be submitted / attached with tender document.
6	Any room renovation which may be required, it is mandatory to conduct site visit (if applicable)
7	Samples should be submitted together with tender or within fourteen (14 days) of the tender closing dates (if applicable).
8	DELIVERY PERIOD: (Please state) Not More Than 90 days upon confirmation
9	PRICE VALIDITY: The quotation shall remain valid for 12 MONTHS from the final date for the submission of the quotation and no supplier may withdraw his/her quotation within that period. The Government reserves the right to extend this period if deemed necessary provided that such extension to the quotation validity period shall have written consent of the supplier(s).
10	The equipment supplied must be newly manufactured, unused, and in its original, sealed packaging . The equipment must not be previously owned, refurbished, or reconditioned in any form. During delivery , the vendor is required to provide proof of manufacture date confirming the equipment is new .

SECTION 3

TENDER FORM

To:

TENDER REFERENCE NO.: KK/253/2026/UPP

INVITATION TO TENDER

SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING OF COLPOSCOPY SYSTEM FOR OBSTETRIC & GYNECOLOGY, WOMEN AND CHILDREN CENTRE RAJA ISTERI PENGIRAN ANAK SALEHA HOSPITAL

TENDER OF (*name of tenderer*) _____

Company/Business Registration No.: _____

Tender Closing Date: _____

SCOPE OF WORK			
Please ☒ Tick where appropriate	Yes	No	Remarks
Supply of <u>one (1) set</u> of <u>Colposcopy</u> for Obstetrics and Gynaecology Department, RIPASH			

SECTION 1 – USER REQUIREMENTS				
1	EQUIPMENT SPECIFICATION			
Please ☒ Tick where appropriate		Yes	No	Remarks
1.1	The system shall be a high-definition digital colposcopy system designed for examination, diagnosis, documentation, and minor procedures of the cervix, vagina, and vulva in gynaecology clinics and operating theatres.			
1.2	The system shall be suitable for use in gynaecology clinic and minor procedure room			
1.3	System configuration			

	The system shall consist of, but not limited to: - Colposcope/imaging unit - Illumination system - Camera/imaging system - Medical-grade display monitor - Mobile stand - Image documentation software.			
1.4	Configuration: Floor stand, Trolley Mounted with lockable wheels			
1.5	Compact endoscopy system			
1.5.1	Configuration: Trolley Mounted with lockable wheels and drawer			
1.5.2	Display			
1.5.2.1	Medical-grade monitor with high-definition display			
1.5.2.2	Monitor display size: Minimum 18" or better			
1.5.2.3	Touch-screen or physical control interface acceptable			
1.5.3	Optical and imaging system			
1.5.3.1	Integrated image processing module			
1.5.3.2	Full HD camera head for full HD imaging resolution or higher			
1.5.3.3	High-quality optical system with clear image reproduction and colour accuracy.			
1.5.4	Integrated LED light source with adjustable light intensity control			
1.5.5	Documentation and connectivity			
1.5.5.1	Image and Video Output: Digital, with USB export capability			
1.5.5.2	Complete with Interface for storing and exporting video files			
1.5.5.3	Monitor supports data export via USB for clinical documentation			
1.6	Telescope for colposcopy procedure			
1.6.1	Autoclavable 90° telescope for colposcopy procedures with integrated illumination, or equivalent technology.			

1.6.2	Working distance: Approximately 25-75cm or better			
1.6.3	Magnification: 7.5x to 15x or better			
1.6.4	Compatible with the compact endoscopy system offered			
1.7	The system must be DICOM and PACS ready and enabled for future integration with Hospital Information System (HIS) where required.			
1.8	The system shall comply with relevant international medical device safety standards (e.g IEC 60601-1 or equivalent).			
1.9	Consumables and Accessories: The Tenderer shall supply the required accessories and consumables to ensure full functionality of the machine, including but not limited to the following: - Cover - Suitable sterilization and storage containers shall be supplied for all reusable components of the system requiring sterilization.			

2. END USER AND TECHNICAL TRAINING				
Please ☒ Tick where appropriate		Yes	No	Remarks
2.1	Conduct user training to the all-end users by an application specialist or competent local engineer including but not limited to: • Basic user operation, user troubleshooting and user maintenance • Provide Operating manual (Hardcopy and/or Softcopy)			
2.1.1	Tenderer must prepare a training attendance or proof of training done to end user during commissioning and the refresher course (6) months after commissioning.			
2.2	Introductory Technical Training to Biomedical Engineers and Technicians at BME Office by competent Tenderer's Engineer/Technicians that includes but not limited to: • Troubleshooting and basic corrective maintenance • Handling and basic inspection maintenance *(Two sessions/groups if required)			

3. WARRANTY				
Please ☒ Tick where appropriate		Yes	No	Remarks
3.1	Tenderer to include warranty period of at least TWO (2) years			
3.2	Tenderers to ACKNOWLEDGE the Warranty Undertaking Form in Section 4 stating the terms of warranty provided for the equipment in the tender for the period of two years. This includes but not limited to: • Scope of Warranty • One-time Planned Preventive Maintenance Per Year during warranty • Comprehensive Corrective Maintenance of Main Unit			

4. PRODUCT DEMONSTRATION (UPON REQUIREMENT)				
Please ☒ Tick where appropriate		Yes	No	Remarks
7.1	If requested by the Government, vendor shall provide a product demonstration as part of the evaluation process The tenderer shall provide either of the following demonstration modes: - Physical Demonstration (In-Person): A complete working unit must be brought to the designated demonstration site as specified by the procuring entity. The demonstration shall be conducted by qualified personnel from the supplier. - Online Demonstration (Virtual): The demonstration may be conducted via a live video conferencing platform (e.g., Zoom, MS Teams). The session must include: - Live operation of the actual product. - Real-time interaction to address questions or perform requested functions. - High-definition video and clear audio for full visibility and understanding.			

SECTION 2 – PRICE PROPOSAL		
PURCHASE PRICE	PER UNIT	BND\$
	TOTAL	BND\$

SECTION 3 - PROCUREMENT AND TECHNICAL SPECIFICATION			
BRAND:		MODEL:	
COUNTRY OF ORIGIN:		YEAR INTRODUCED TO MARKET:	
WARRANTY PERIOD:		LAST COUNTRY SOLD TO:	
PRICE VALIDITY: [AT LEAST <u>ONE (1) YEAR</u> PRICE VALIDTY]		DELIVERY TIME:	

SECTION 3 - PROCUREMENT AND TECHNICAL SPECIFICATION					
AUTHORIZED DISTRIBUTOR: (AUTHORIZED DISTRIBUTOR LETTER ATTACHED)	APPOINTED BRUNEI DISTRIBUTOR				
		PROCURE FROM OVERSEA AUTHORIZED DISTRIBUTOR	COMPANY NAME:		
			COMPANY ORIGIN:		
DETAILED BROCHURE INCLUDED		YES		NO	☒ or specify where appropriate
USER AND SERVICE MANUALS:		YES		NO	Tenderers to acknowledge that they must provide at least TWO sets of USER AND SERVICE manuals when applying commissioning form. One Set for End User, One Set for BME. (Please provide hardcopy or softcopy)
MAINS POWER SUPPLY:		220V-240V		OTHERS:	
		50-60HZ		OTHERS:	
BATTERY		RECHARGEABLE			SINGLE-USE
		OTHERS:			
	TYPE OF BATTERY:				
	RATING:				
POWER ADAPTER/CHARGER OUTPUT RATING:					
EQUIPMENT AMBIENT OPERATING TEMPERATURE RANGE:					
INTERNATIONAL SAFETY STANDARD Must comply to at least 1 safety Standards and certification (Please attached the copy of stated standards and certifications)					☒ Tick where appropriate ☐ US FDA Standard, ☐ European Union CE MARK, ☐ Australian TGA Standard, ☐ Canadian CSA Standard or ☐ Japanese JIS Standard. Others (Please specify): _____
NUMBER OF TECHNICAL SUPPORT (ENGINEER/TECHNICIAN) Please provide training or certification for locals who is trained/certified	LOCAL			☐ Trained / Certified ☐ Not yet trained on the product	
	OVERSEA (SPECIFY LOCATION)		NEAREST LOCATION:		
DIMENSIONS AND WEIGHT OF MAIN UNIT:		☐ mm ☐ cm ☐ inch		☐ Kilogram (Kg) ☐ Gram(g) ☐ Pound (lbs)	
EQUIPMENT WHOLE LIFE TIME SUPPORT:	The supplier shall ensure that spare parts for the equipment are available for a minimum of 10 years after installation, with the support period extending beyond the expected lifecycle of the equipment. No of years: _____ (Please specify)				

SECTION 4 – WARRANTY UNDERTAKING FORM

Tenderer, on behalf of the manufacturer, acknowledged and agrees that when equipment is under the warranty period, must cover the scope of normal warranty below at no additional cost:

NORMAL WARRANTY

- Warrants the supplied medical equipment and its accessories to be in good condition, in working order and free from defects to the extent such equipment do not comply with specifications, under normal use for the warranty period. The scope of warranty covers to its maximum extent permitted by applicable law.
- During warranty, tenderer must rectify issues arise from any mechanical, technical or software faulty as soon as it is reported.
- **Exchange warranty;** Providing replacement units:
 - A. Warranty against defects – Manufacturing defects or Equipment malfunction resulted from mechanical, electrical or software failure during Commissioning or within the first _____ months of use
 - B. Faulty workmanship or unsatisfactory condition during delivery or commissioning
 - C. If a unit or accessory is deemed used item or refurbished item (not a new unit) by the user and BME Unit.
- **Two time Planned Preventive Maintenance (PPM) PER YEAR** according to Manufacturer's Preventive Maintenance Guideline and to include one-time replacements of battery at the end of 2 years warranty period or any other relevant parts to prolong equipment lifespan.

EXCLUSION FROM WARRANTY

MOH understand that the following circumstances are not covered in the warranty and Tenderer may quote for repair and subject to MOH approval:

- Unauthorized modifications - an alteration or repair by anyone other than the Manufacturer or Authorized agent during warranty period.
- Accidental damage or problems caused by negligence or mishandling, subject to appropriate justification by both parties.
- Vandalism and Natural disasters
- Normal wear and tear

ANY OTHER EXCLUSION

Tenderer may propose below to include items or terms which is not listed in the exclusion list above for MOH consideration.

TENDERER ACKNOWLEDGMENT

COMPANY CHOP AND SIGNATURE

	TERMS AND CONDITIONS	VENDOR'S OFFER (PLEASE STATE)
1	Tenderer must be registered with the Ministry of Health.	
2	TENDER FORM should be filled completely including the USER REQUIREMENT FORM (if available). Submission of incomplete form <u>MAY</u> cause DISQUALIFICATION OF TENDER .	
3	Each tenderer is allowed to quote ONE BRAND WITH ONE PRICE ONLY for each item. Submission of more than one brand and price will cause DISQUALIFICATION OF TENDER .	
4	All consumables supplied throughout this tender <u>shall</u> have a minimum expiry date of twelve (12) months / on delivery (if applicable). Should the consumables be urgently needed, provision of consumables with expiry date of less than twelve (12) months should be first agreed by the User before delivery is made (if applicable).	
5	Brochures / catalogues should be submitted / attached with tender document.	
6	Any room renovation which may be required, it is mandatory to conduct site visit (if applicable)	
7	Samples should be submitted together with tender or within fourteen (14 days) of the tender closing dates (if applicable).	
8	DELIVERY PERIOD: (Please state) Not More Than 90 days upon confirmation	
9	PRICE VALIDITY: The quotation shall remain valid for 12 MONTHS from the final date for the submission of the quotation and no supplier may withdraw his/her quotation within that period. The Government reserves the right to extend this period if deemed necessary provided that such extension to the quotation validity period shall have written consent of the supplier(s).	
10	The equipment supplied must be newly manufactured, unused, and in its original, sealed packaging . The equipment must not be previously owned, refurbished, or reconditioned in any form. During delivery , the vendor is required to provide proof of manufacture date confirming the equipment is new .	

1. We offer and undertake on your acceptance of our Tender to supply and deliver the above mentioned goods in accordance with your Invitation To Tender.
2. Our Tender is fully consistent with and does not contradict or derogate from anything in your Invitation To Tender. We have not qualified or changed any of the provisions of your Invitation To Tender.
3. We shall execute a formal agreement in the appropriate form set out in Section 4 – Contract of the Invitation to Tender together with such further terms and conditions, if any, agreed between the Government and us.
4. OUR OFFER IS VALID FOR **TWELVE (12)** CALENDER MONTHS FROM THE TENDER CLOSING DATE.
5. When requested by you, we shall extend the validity of this offer.
6. We further undertake to give you any further information which you may require.

Dated this day of 20 .

Signature of authorised officer of Tenderer
Name:
Designation:

Tenderer's official stamp: