FORM TO BE USED

FORM B

UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM

SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING TEN (10) UNIT OF DENTAL AIR COMPRESSOR FOR DEPARTMENT OF DENTAL SERVICES (NON-CLUSTERING)

NO.	TERMS AND CONDITIONS	VENDOR'S OFFER (PLEASE STATE)
1	Tenderer must be registered with the Ministry of Health.	
2	**TENDER FORM should be filled** completely including the **USER REQUIREMENT FORM** (if available). Submission of incomplete form **MAY** cause **DISQUALIFICATION OF TENDER.**	
3	Each tenderer is allowed to quote **ONE BRAND WITH ONE PRICE ONLY** for each item. Submission of more than one brand and price will cause **DISQUALIFICATION OF TENDER.**	
4	All consumables supplied throughout this tender shall have a minimum expiry date of **twelve (12) months / on delivery**. Should the consumables be urgently needed, provision of consumables with expiry date of less than twelve (12) months should be first agreed by the User before delivery is made.	
5	**Brochures / catalogues should be submitted / attached** with tender document.	
6	Any **room renovation** which may be required, **it is mandatory to conduct site visit** (if applicable)	
7	**DELIVERY PERIOD:** (Please state) Between 90 – 120 days upon confirmation	**(Yes / No)** **(If No, please specify)**
8	**PRICE VALIDITY:** The quotation shall remain valid for **12 MONTHS** from the final date for the submission of the quotation and no supplier may withdraw his/her quotation within that period. The Government reserves the right to extend this period if deemed necessary provided that such extension to the quotation validity period shall have written consent of the supplier(s).	

UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM

SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING TEN (10) UNIT OF DENTAL AIR COMPRESSOR FOR DEPARTMENT OF DENTAL SERVICES (NON-CLUSTERING)

Location for installation	**Item 1: Single Cylinder** 1. Cubicle 1, Block A, National Dental Centre 2. Room 10, Block B, National Dental Centre 3. Cubicle 5, Block B, National Dental Centre 4. Room 51, Muara Health Clinic 5. Sekolah Rendah Pulaie 6. Sekolah Rendah Bukit Beruang, Tutong 7. Sekolah Rendah Penaga **Item 2: Dual Cylinder** 1. Room 2, Sungai Liang Health Clinic 2. Room 5, PMMPMHAMB Hospital, Tutong 3. Room 28, Pengkalan Batu Health Clinic

SECTION I - USER REQUIREMENT		YES	NO	PLEASE SPECIFY THE SPECIFICATION/ SCOPE OFFERED FAILURE TO FILL THE OFFER DETAIL SHALL BE IDENTIFIED AS NON-COMPLIANCE TO THE OFFER
Item 1: Dental Air Compressor: Single Cylinder				
Standard features	For dentistry purpose.			
	With membrane-drying unit; To prevent formation of moisture as to protect instruments and ensures the hygiene.			
	Oil-free, dry, hygienic and odourless compressed air			
	Suitable for continuous operation			
	Compatible with the current dental chair system at Ministry of Health			
No. of cylinder	1 unit			
Flow rate at 5 bar	Approximately 70 litres/min			
Tank capacity	Not less than 20 litres			
Pressure range	Min: Not more than 6 bar Max: Not less than 7.8 bar			
Noise level (without cover)	Not more than 64 dB			

UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM

SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING TEN (10) UNIT OF DENTAL AIR COMPRESSOR FOR DEPARTMENT OF DENTAL SERVICES (NON-CLUSTERING)

SECTION I - USER REQUIREMENT		YES	NO	PLEASE SPECIFY THE SPECIFICATION/ SCOPE OFFERED FAILURE TO FILL THE OFFER DETAIL SHALL BE IDENTIFIED AS NON-COMPLIANCE TO THE OFFER
Item 1: Dental Air Compressor: Single Cylinder				
Air filter system	At least 3 stage air filtering system for hygienic compressed air and greater reliability which are as the following:			
	• Air intake filter			
	• Bacteria filter; A high performance particulate air filter			
	• Coalescence filter with water collecting container			
Accessories:	**Accessories to be included (The Equipment Should Come With All Necessary Accessories For It To Be Fully Operational)**			

UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM

SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING TEN (10) UNIT OF DENTAL AIR COMPRESSOR FOR DEPARTMENT OF DENTAL SERVICES (NON-CLUSTERING)

SECTION I - USER REQUIREMENT		YES	NO	PLEASE SPECIFY THE SPECIFICATION/ SCOPE OFFERED FAILURE TO FILL THE OFFER DETAIL SHALL BE IDENTIFIED AS NON-COMPLIANCE TO THE OFFER
Item 2: Dental Air Compressor: Dual Cylinder				
Standard features	For dentistry purpose.			
	With membrane-drying unit; To prevent formation of moisture as to protect instruments and ensures the hygiene.			
	Oil-free, dry, hygienic and odourless compressed air			
	Suitable for continuous operation			
	Compatible with the current dental chair system at Ministry of Health			
No. of cylinder	2 units			
Flow rate at 5 bar	Approximately 137 litres/min			
Tank capacity	Not less than 20 litres			
Pressure range	Min: Not more than 6 bar Max: Not less than 7.8 bar			
Noise level (without cover)	Not more than 68 dB			
Air filter system	At least 3 stage air filtering system for hygienic compressed air and greater reliability which are as the following:			
	• Air intake filter			
	• Bacteria filter; A high performance particulate air filter			
	• Coalescence filter with water collecting container			
Accessories:	**Accessories to be included (The Equipment Should Come With All Necessary Accessories For It To Be Fully Operational)**			

UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM

SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING TEN (10) UNIT OF DENTAL AIR COMPRESSOR FOR DEPARTMENT OF DENTAL SERVICES (NON-CLUSTERING)

SECTION II - PROCUREMENT REQUIREMENT	
Model and Brand of Equipment.	
Country of Origin.	
Unit (B$)	
Total Price Offered (CIF)	
Country of Origin	
Year manufactured.	
Warranty period (at least 1 year)	_____ years
Delivery period	_____ weeks
Price validity (not less than 12 months)	_____ years

UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM

SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING TEN (10) UNIT OF DENTAL AIR COMPRESSOR FOR DEPARTMENT OF DENTAL SERVICES (NON-CLUSTERING)

TECHNICAL SPECIFICATIONS/REQUIREMENTS	
Mains Supply 240VAC 50 Hz.	
State equipment ambient operating temperature range.	
The equipment must comply to either one of the five international safety standards, namely; US FDA Standard, European Union CE MARK, Australian TGA Standard, Canadian CSA Standard or Japanese JIS Standard.	
State number of locally stationed engineers/technicians and the nearest overseas support.	
State overall dimensions of equipment.	
State net weight of the equipment (kg)	
Equipment whole life time support Supplier must guarantee spare parts are available at least for **10 years** after the installation of the equipment.	_____years
Brochure Tenderers must submit with their offer, a detailed brochure which supports the information the tenderer has entered in the "form-to-be-used (FORM B)" for reference.	
User and Service Manuals Two hard copies and one soft copy [PDF FORMAT] to be submitted to BME office during commissioning.	
Spare-parts and consumables Comprehensive spare-parts and consumables listings in the form of softcopy must be provided. Listing must have the following details; Part Number and Part Description.	
Warranty planned preventive maintenance. Minimum of two times PPM per year during the equipment warranty, six months after the date of commissioning and at the end of warranty period.	

UNIT TAWARAN
KEMENTERIAN KESIHATAN
NEGARA BRUNEI DARUSSALAM

SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING TEN (10) UNIT OF DENTAL AIR COMPRESSOR FOR DEPARTMENT OF DENTAL SERVICES (NON-CLUSTERING)

TECHNICAL SPECIFICATIONS/REQUIREMENTS	

Training
Tenderers are required to conduct technical training to Biomedical Engineers and Technicians.
 	
• Training to be conducted locally, tenderers are required to:
  • Provide training materials, test equipment, demo equipments, etc.
  • Provide training to two groups of technical staffs.
  • Provide 2 days (minimum) of training for each group.
  • Training to be conducted at BME office/ respective health clinic	