LAMPIRAN - A1

RUJ: BME/A2K/(1778)/2026/HTD (A20) **24/10/2026**

REPAIR THREE (3) UNITS OF DYNAFLOW FUMEHOODS AT DEPARTMENT OF SCIENTIFIC SERVICES

SCOPE OF WORK

Repair services for the following three (3) units of Dynaflow fumehoods at Department of Scientific Services:

1. BME11609, S/N: 14005/14 at Environment laboratory
2. BME0002437, S/N: 15591/2 at Food Chemistry laboratory
3. BME0011601, S/N:14005/11 at Food Chemistry laboratory

- Scope of work to follow equipment manual/recommendation specific for this system which include the following where necessary:

(i) installation of items listed in spareparts list in Lampiran A2
  - Item 1 for all three units
  - Item 2 for BME0002437 only

(i) alignment
(iii) calibration
(iv) optimisation
(vi) any other related maintenance works required after installation of items listed in spareparts list

- Functional check post-maintenance.

- Disposal of any used spare parts/pm kits, unless otherwise instructed.

- Work to be carried out by engineer(s) authorised in providing maintenance to the equipment.

LAMPIRAN - A2

RUJ: BME/A2K/(1778)/2026/HTD (A20)

REPAIR THREE (3) UNITS OF DYNAFLOW FUMEHOODS AT DEPARTMENT OF SCIENTIFIC SERVICES

	SPAREPART	PART NUMBER	QUANTITY
1	MK5 main controller board with keyboard		3
2	Variable speed controller 0.75kw		1