LAMPIRAN - A1

RUJ: BME/A2K/(1780)/2026/HTD (B8) **31/10/2026**

ONE OFF PPM FOR TWO (2) UNITS OF HEAT SEALER (STERILE POUCH), BRAND: GETINGE, MODEL: 6003000195 AT CSSD RIPAS HOSPITAL.

SCOPE OF WORK

One off PPM for two (2) units of Heat sealer (sterile pouch), Brand: Getinge at CSSD Ripas Hospital. Detail are as follow:

1. BME26018, Brand: Getinge, Model: 6003000195, SN: PPR20210056, at CSSD W/order 2026-13818
2. BME16019, Brand: Getinge, Model: 6003000195, SN: PPR20210059, at CSSD W/order 2026-14397

- Scope of work to follow maintenance manual/recommendation specific for models, which include the following where necessary:
(i) inspection
(ii) cleaning
(iii) installation of PM kits, consumables and spare parts
(iv) alignment
(v) calibration
(vi) any other related preventive maintenance works required

- Functional check post-maintenance.

- Apply preventive maintenance sticker/tag after successful service.

- Disposal of any used spare parts, unless otherwise instructed.

- The quotation submitted shall remain valid for a period of not less than twelve (12) months from the date of submission.

- Work on the system is to be carried out by engineer(s) authorised in providing maintenance to the equipment.

- Vendor(s) to submit maintenance checklist/details/scope specific for the equipment along with other necessary documents to join this tender.

LAMPIRAN - A2

RUJ: BME/A2K/(1780)/2026/HTD (B8)

ONE OFF PPM FOR TWO (2) UNITS OF HEAT SEALER (STERILE POUCH), BRAND: GETINGE, MODEL: 6003000195 AT CSSD RIPAS HOSPITAL.

	SPAREPART	PART NUMBER	QUANTITY
1	PM KIT PROSEAL V-6	6005500597	2