LAMPIRAN - A2

RUJ: BME/A2K/(1790)/2026/HTD (B4)

ONE OFF PLANNED PREVENTIVE MAINTENANCE FOR HOLTERS AND ABPM AT CARDIO CLINIC, RIPASH

SPAREPART	PART NUMBER	QUANTITY

LAMPIRAN - A1

RUJ: BME/A2K/(1790)/2026/HTD (B4) **10/11/2026**

ONE OFF PLANNED PREVENTIVE MAINTENANCE FOR HOLTERS AND ABPM AT CARDIO CLINIC, RIPASH

SCOPE OF WORK

One off preventive maintenance service for TWELVE (12) units of HOLTERS (OCT) and SIX (6) units of ABPM at CARDIOLOGY, RIPASH. The details of the equipment are as follows:
HOLTERS, BRAND : SCHILLER, MODEL : MEDILOG AR
1. BME31704,S/N: 3060.016927
2, BME31705, S/N: 3060.016763
3. BME31706,S/N: 3060.016739
4. BME31707, S/N: 3060.016756
5. BME31708, S/N: 3060.016742
6. BME31709, S/N: 3060.022573
7. BME31710, S/N: 3060.022589
8, BME31711, S/N: 3060.022595
9. BME31712, S/N: 3060.022577
10. BME31713, S/N: 3060.022660
11. BME31714, S/N: 3060.022683
12. BME31715, S/N: 3060.022534

ABPM, BRAND : SCHILLER, MODEL : BR 102+
1. BME31716,S/N: 293.22131
2, BME31717, S/N: 293.22132
3. BME31718,S/N: 293.22133
4. BME31719, S/N: 293.22134
5. BME31720, S/N: 293.2137
6. BME31721, S/N : 293.22137

- Scope of work to follow maintenance manual/recommendation specific for models, which include the following where necessary:
(i) visual inspection
(ii) cleaning
(iii) functional test
(iv) performance test
(v) battery inspection
(vi) Apply preventive maintenance sticker/tag after successful service
(vii) any other related preventive maintenance works required