TENDER REF. NO.: KK/264/2026/UPP(TC)

MINISTRY OF HEALTH
NEGARA BRUNEI DARUSSALAM

TO SUPPLY AND DELIVERY OF MEDICAL CONSUMABLES FOR HAEMATOLOGY UNIT, RAJA ISTERI PENGIRAN ANAK SALEHA (RIPAS) HOSPITAL, MINISTRY OF HEALTH FOR A PERIOD OF THREE (3) YEARS USAGE

TENDER FEE : $500.00

RECEIPT NO. :

CLOSING DATE : ON TUESDAY, 01st December 2026

TIME : 2.00 PM

FOA :

THE CHAIRMAN
MINI TENDER BOARD, TENDER BOX
GROUND FLOOR, MINISTRY OF HEALTH
COMMONWEALTH DRIVE
BANDAR SERI BEGAWAN BB3910
NEGARA BRUNEI DARUSSALAM

[CLUSTERING]

SECTION 2

SPECIFICATIONS

TENDER REFERENCE NO: KK/264/2026/UPP(TC)

INVITATION TO TENDER
TO SUPPLY AND DELIVERY OF MEDICAL CONSUMABLES FOR HAEMATOLOGY UNIT, RAJA ISTERI PENGIRAN ANAK SALEHA (RIPAS) HOSPITAL, MINISTRY OF HEALTH FOR A PERIOD OF THREE (3) YEARS USAGE

	USER'S REQUIREMENTS		
NO	**ITEM DESCRIPTIONS AND SPECIFICATIONS**	**PACKAGING SIZE**	**TOTAL ESTIMATE USAGE / YEAR**
1	**Trima Accel® LRS® Platelet, Plasma Set** • **AC Filter, Replacement Fluid Filter, Additive Solution Filter:** Exist to maintain functionally closed system • **Extracorporeal Blood Volume: Approx.** 196mL • **Needle Gauge:** 17 gauge • **Return Reservoir Filter:** Prevents particles (200 microns or larger) from entering the line at the bottom of the reservoir, and return reservoir prevents air from entering the return line • **Platelet Storage Bag Volume:** 100mL to 400mL • Platelet Concentration: 0.7 to 2.1 x 10^6/µL • **Maximum Total Platelets per Bag (Platelets in Plasma):** ≤ 5.1 × 10^{11} platelets • **Platelet Storage Temperature:** 22°C ± 2 °C • **Plasma Storage Bag Volume:** If frozen: Up to 600mL; If liquid: Up to 1,000mL • Disposable can be used upwards to 24 hours after disposable set has been opened and set loaded, if loaded and primed must be used within business day	6 units/box	134 boxes

	USER'S REQUIREMENTS		
NO	**ITEM DESCRIPTIONS AND SPECIFICATIONS**	**PACKAGING SIZE**	**TOTAL ESTIMATE USAGE / YEAR**
2	**Anticoagulant Citrate Dextrose Solution USP(ACP) Solution A, 750ml** • Sterile, non-pyrogenic • Single-use container • For use only with apheresis devices offered • Not for direct intravenous infusion • Store up to 25°C (see USP controlled room temperature) • Each 100mL contains: 2.45g Dextrose Monohydrate USP, 2.20g Sodium Citrate Dihydrate USP, 0.80g Citric Acid Monohydrate USP, In Water for Injection USP	12 units/box	75 boxes
3	**Spectra Optia™ Exchange Set** • **Procedure:** Therapeutic Plasma Exchange (TPE) and Red Blood Cell (RBC) Exchange • **Remove Bag Volume:** 6L • **Extracorporeal Blood Volume:** Approx. 185mL • **Inlet Line Trap:** Traps large particles • **Reservoir Volume:** (a) Between the high- and low-level sensors: 49mL; (b) Entire Reservoir: 56mL; (c) Reservoir prevents more than 0.5mL of air from entering the return line • **Reservoir Filter:** Prevents particles (200 microns or larger) from entering the return line exiting the reservoir • **AC Check Valve:** Prevents the free flow of AC through the tubing set • **RBC Residual Volume:** (a) When rinseback is performed (average volume): Exchange Set - 10mL (b) When rinseback is not performed: Equal to the ECV of the tubing set multiplied by the patient's Hct • **Channel Life:** 6 hours	6 units/box	10 boxes

	USER'S REQUIREMENTS		
NO	**ITEM DESCRIPTIONS AND SPECIFICATIONS**	**PACKAGING SIZE**	**TOTAL ESTIMATE USAGE / YEAR**
	• Disposable can be used upwards to 24 hours after disposable set has been opened and set loaded, if loaded and primed must be used within business day		
4	**Spectra Optia™ IDL Set** • **Procedure:** For Continuous Mononuclear Cell Collection (CMNC), White Blood Cell Depletion (WBCD) and Platelet Depletion (PD) • **Extracorporeal Blood Volume:** Approx. 297mL • **Collection Bag Volume:** (a) 940mL, if bag is not centrifuged (b) 400mL, if bag is centrifuged (Collection bag may be centrifuged for up to 5000g) • **Plasma Bag Volume:** (a) Liquid: 800mL; (b) Frozen: 600mL •**Reservoir Volume:** (a) Between the high- and low-level sensors: 49mL; (b) Entire Reservoir: 56mL; Reservoir prevents more than 0.5mL of air from entering the return line • **Reservoir Filter:** Prevents particles (200 microns or larger) from entering the return line exiting the reservoir • **Sterile Barrier Filter of Collection and IDL Set:** 0.2 microns, and helps to maintain a functionally closed system • **Inlet Line Trap:** Traps large particles • **AC Check Valve:** Prevents the free flow of AC through the tubing set • **RBC Residual Volume:** (a) When rinseback Set - 17mL (b) When rinseback is not performed: Equal to the ECV of the tubing set multiplied by the patient's Hct • **Channel Life:** 8 hours	6 units/box	2 boxes

	USER'S REQUIREMENTS		
NO	**ITEM DESCRIPTIONS AND SPECIFICATIONS**	**PACKAGING SIZE**	**TOTAL ESTIMATE USAGE / YEAR**
	• Disposable can be used upwards to 24 hours after disposable set has been opened and set loaded, if loaded and primed must be used within business day **Procedure: For Bone Marrow Processing Only** • **BMP Bag Volume:** 3L • **Administration line filter on BMP Accessory Set**: 200 microns, used for additional filtration of bone marrow		

Section/Unit	Haematology Unit RIPAS Hospital	
Section/Unit Ref No.:		
Person to Contact	Name : NORIMAH BINTI YUNUS	
	E-mail : norimah.yunus@moh.gov.bn	
	Tel.No. : 2242424 EXT 5255 / 5222	Fax No. :
FOR ADMINISTRATION USE ONLY		
Ref.No.		
Advertisement Ref. No.		Date :

NO	TERMS AND CONDITIONS
1.0	Tenderer must be registered with the Ministry of Health.
2.0	**TENDER FORM should be filled** completely including the **USER REQUIREMENT FORM** (if available). Submission of incomplete form **MAY** cause **DISQUALIFICATION OF TENDER.**
3.0	Each tenderer is allowed to quote **ONE BRAND WITH ONE PRICE ONLY** for each item. Submission of more than one brand and price will cause **DISQUALIFICATION OF TENDER.**
4.0	All consumables supplied throughout this tender shall have a minimum expiry date of twelve (12) months / on delivery. Should the consumbles be urgently needed, provision of consumbles with expiry date of less than twelve (12) months should be first agreed by the User before delivery is made.
5.0	**Brochures / catalogues should be submitted / attached with tender document.**
6.0	**SERVICE AND AFTER SALES SUPPORT**
6.1	All consumables supplied throughout this tender shall have a minimum expiry date of **six (6) months on delivery**. Should the consumable be urgently needed, provision of a consumable with expiry date of less than six (6) months should be first agreed by the User of the particular laboratory before delivery is made.
6.2	Letter of Undertaking (LOU) shall be produced upon each delivery of consumable with expiry date of less than **six (6)** months and vendor shall declare in the LOU that unused, unopened, expired kits will be replaced accordingly.
6.3	User shall have the rights to return any items, and to be replaced at no extra cost, if found not meeting the acceptance criteria upon opening a pack such as, but not limited to, the following: 1. Tampered or damaged packaging 2. Evident of leakage or damaged products 3. Expired products that are evidently less than the requirement mentioned in para 4.1 calculated from delivery date 4. Leakage upon delivery
6.4	Vendor shall submit samples of the offered items directly to the Users no later than 7 days after the Closing Date of this advertisement or as required by the Users.
7.0	EXIT CLAUSE: The tender contract shall be automatically terminated even though tender has not yet expired and this shall be in effect due to, but not limited to, the following: **1.** When the testing is no longer required or relevant i.e. test is obsolete, to the laboratory or department. **2.** When the item(s) set out in this tender is/are no longer required by the laboratory or the Department. **3.** When the approved budget allocation for this tender contract has been used up before the tender contract expires whereby a renewal of tender shall be submitted by the user for an open advertisement subject to approval by the Mini Tender Board (*Lembaga Tawaran Kecil*).
8.0	DELIVERY PERIOD: **Preferably 4 – 8 weeks and no later than 12 weeks after issue of Purchase Order**
9.0	PRICE VALIDITY: The quotation shall remain valid for **12 months** from the final date for the submission of the quotation and no supplier may withdraw his/her quotation within that period. The Government reserves the right to extend this period if deemed necessary provided that such extension to the quotation validity period shall have written consent of the supplier(s).

SECTION 3

FORMS TO BE USED

CONTENTS

SCHEDULE 1 – TENDER FORM ……..……..……1

SCHEDULE 2 – INFORMATION SUMMARY ..……..……..…4

SCHEDULE 3 –SUB-CONTRACTS ……....……..……..……5

SCHEDULE 4 –COMPANY BACKGROUND ……..……..6

SCHEDULE 5 –REFERENCES ……..……..……..……..7

SCHEDULE 6 –SUBMISSION OF SAMPLE ... ……..……...8

SCHEDULE 7 - LETTER OF DECLARATION ……11

SCHEDULE 1

TENDER FORM

To:

TENDER REFERENCE NO: KK/264/2026/UPP(TC)

INVITATION TO TENDER
TO SUPPLY AND DELIVERY OF MEDICAL CONSUMABLES FOR HAEMATOLOGY UNIT, RAJA ISTERI PENGIRAN ANAK SALEHA (RIPAS) HOSPITAL, MINISTRY OF HEALTH FOR A PERIOD OF THREE (3) YEARS USAGE

TENDER OF (*name of tenderer*) _____

Company/Business Registration No _____

Tender Closing Date: _____

1. SUPPLY OF REAGENTS

1.1 To supply reagents and associated consumables (calibrators, controls, accessories and consumables) for the tests listed below.

APPENDIX A: SUMMARY OF UNIT PRICE OF REAGENT KIT
(To be completed by Vendor for submission)

	USER'S REQUIREMENTS			**VENDOR'S OFFER**					
NO	**ITEM DESCRIPTIONS AND SPECIFICATIONS**	**PACKAGING SIZE**	**TOTAL ESTIMATE USAGE / YEAR**	**ITEM DESCRIPTIONS AND SPECIFICATIONS**	**PART/ CATALOGUE NUMBER AND BRAND**	**PACKAGING SIZE**	**TOTAL QUANTITY OFFERED / YEAR**	***COST PER UNIT (B$)**	**TOTAL COSTS (B$)**
1	**Trima Accel® LRS® Platelet, Plasma Set** • **AC Filter, Replacement Fluid Filter, Additive Solution Filter:** Exist to maintain functionally closed	6 units/box	134 boxes						

	USER'S REQUIREMENTS			VENDOR'S OFFER					
NO	**ITEM DESCRIPTIONS AND SPECIFICATIONS**	**PACKAGING SIZE**	**TOTAL ESTIMATE USAGE / YEAR**	**ITEM DESCRIPTIONS AND SPECIFICATIONS**	**PART/ CATALOGUE NUMBER AND BRAND**	**PACKAGING SIZE**	**TOTAL QUANTITY OFFERED / YEAR**	***COST PER UNIT (B$)**	**TOTAL COSTS (B$)**
	system • **Extracorporeal Blood Volume: Approx.** 196mL • **Needle Gauge:** 17 gauge • **Return Reservoir Filter:** Prevents particles (200 microns or larger) from entering the line at the bottom of the reservoir, and return reservoir prevents air from entering the return line • **Platelet Storage Bag Volume:** 100mL to 400mL • Platelet Concentration: 0.7 to 2.1 x 10^6/µL • **Maximum Total Platelets per Bag (Platelets in Plasma):** ≤ 5.1 × 10^{11} platelets • **Platelet Storage Temperature:** 22°C ± 2 °C • **Plasma Storage Bag Volume:** If frozen: Up to 600mL; If liquid: Up to								

	USER'S REQUIREMENTS			VENDOR'S OFFER					
NO	**ITEM DESCRIPTIONS AND SPECIFICATIONS**	**PACKAGING SIZE**	**TOTAL ESTIMATE USAGE / YEAR**	**ITEM DESCRIPTIONS AND SPECIFICATIONS**	**PART/ CATALOGUE NUMBER AND BRAND**	**PACKAGING SIZE**	**TOTAL QUANTITY OFFERED / YEAR**	***COST PER UNIT (B$)**	**TOTAL COSTS (B$)**
	1,000mL • Disposable can be used upwards to 24 hours after disposable set has been opened and set loaded, if loaded and primed must be used within business day								
2	**Anticoagulant Citrate Dextrose Solution USP(ACP) Solution A, 750ml** • Sterile, non-pyrogenic • Single-use container • For use only with apheresis devices offered • Not for direct intravenous infusion • Store up to 25°C (see USP controlled room temperature) • Each 100mL contains: 2.45g Dextrose Monohydrate USP, 2.20g Sodium Citrate Dihydrate USP, 0.80g Citric Acid Monohydrate USP, In Water for Injection USP	12 units/box	75 boxes						

	USER'S REQUIREMENTS			VENDOR'S OFFER					
NO	**ITEM DESCRIPTIONS AND SPECIFICATIONS**	**PACKAGING SIZE**	**TOTAL ESTIMATE USAGE / YEAR**	**ITEM DESCRIPTIONS AND SPECIFICATIONS**	**PART/ CATALOGUE NUMBER AND BRAND**	**PACKAGING SIZE**	**TOTAL QUANTITY OFFERED / YEAR**	***COST PER UNIT (B$)**	**TOTAL COSTS (B$)**
3	**Spectra Optia™ Exchange Set** • **Procedure:** Therapeutic Plasma Exchange (TPE) and Red Blood Cell (RBC) Exchange • **Remove Bag Volume:** 6L • **Extracorporeal Blood Volume:** Approx. 185mL • **Inlet Line Trap:** Traps large particles • **Reservoir Volume:** (a) Between the high- and low-level sensors: 49mL; (b) Entire Reservoir: 56mL; (c) Reservoir prevents more than 0.5mL of air from entering the return line • **Reservoir Filter:** Prevents particles (200 microns or larger) from entering the return line exiting the reservoir • **AC Check Valve:** Prevents the free flow of AC through the tubing set • **RBC Residual Volume:** (a) When rinseback is	6 units/box	10 boxes						

	USER'S REQUIREMENTS			VENDOR'S OFFER					
NO	**ITEM DESCRIPTIONS AND SPECIFICATIONS**	**PACKAGING SIZE**	**TOTAL ESTIMATE USAGE / YEAR**	**ITEM DESCRIPTIONS AND SPECIFICATIONS**	**PART/ CATALOGUE NUMBER AND BRAND**	**PACKAGING SIZE**	**TOTAL QUANTITY OFFERED / YEAR**	***COST PER UNIT (B$)**	**TOTAL COSTS (B$)**
	performed (average volume): Exchange Set - 10mL (b) When rinseback is not performed: Equal to the ECV of the tubing set multiplied by the patient's Hct • **Channel Life:** 6 hours • Disposable can be used upwards to 24 hours after disposable set has been opened and set loaded, if loaded and primed must be used within business day								
4	**Spectra Optia™ IDL Set** • **Procedure:** For Continuous Mononuclear Cell Collection (CMNC), White Blood Cell Depletion (WBCD) and Platelet Depletion (PD) • **Extracorporeal Blood Volume:** Approx. 297mL • **Collection Bag Volume:** (a) 940mL, if bag is not centrifuged	6 units/box	2 boxes						

	USER'S REQUIREMENTS			VENDOR'S OFFER					
NO	**ITEM DESCRIPTIONS AND SPECIFICATIONS**	**PACKAGING SIZE**	**TOTAL ESTIMATE USAGE / YEAR**	**ITEM DESCRIPTIONS AND SPECIFICATIONS**	**PART/ CATALOGUE NUMBER AND BRAND**	**PACKAGING SIZE**	**TOTAL QUANTITY OFFERED / YEAR**	***COST PER UNIT (B$)**	**TOTAL COSTS (B$)**
	(b) 400mL, if bag is centrifuged (Collection bag may be centrifuged for up to 5000g) • **Plasma Bag Volume:** (a) Liquid: 800mL; (b) Frozen: 600mL •**Reservoir Volume:** (a) Between the high- and low-level sensors: 49mL; (b) Entire Reservoir: 56mL; Reservoir prevents more than 0.5mL of air from entering the return line • **Reservoir Filter:** Prevents particles (200 microns or larger) from entering the return line exiting the reservoir • **Sterile Barrier Filter of Collection and IDL Set:** 0.2 microns, and helps to maintain a functionally closed system • **Inlet Line Trap:** Traps large particles • **AC Check Valve:** Prevents the free flow of								

	USER'S REQUIREMENTS			VENDOR'S OFFER					
NO	**ITEM DESCRIPTIONS AND SPECIFICATIONS**	**PACKAGING SIZE**	**TOTAL ESTIMATE USAGE / YEAR**	**ITEM DESCRIPTIONS AND SPECIFICATIONS**	**PART/ CATALOGUE NUMBER AND BRAND**	**PACKAGING SIZE**	**TOTAL QUANTITY OFFERED / YEAR**	***COST PER UNIT (B$)**	**TOTAL COSTS (B$)**
	AC through the tubing set • **RBC Residual Volume:** (a) When rinseback Set - 17mL (b) When rinseback is not performed: Equal to the ECV of the tubing set multiplied by the patient's Hct • **Channel Life:** 8 hours • Disposable can be used upwards to 24 hours after disposable set has been opened and set loaded, if loaded and primed must be used within business day **Procedure: For Bone Marrow Processing Only** • **BMP Bag Volume:** 3L • **Administration line filter on BMP Accessory Set**: 200 microns, used for additional filtration of bone marrow								

Section/Unit	Haematology Unit RIPAS Hospital	
Section/Unit Ref No.:		
Person to Contact	Name : NORIMAH BINTI YUNUS	
	E-mail : norimah.yunus@moh.gov.bn	
	Tel.No. : 2242424 EXT 5255 / 5222	Fax No. :
FOR ADMINISTRATION USE ONLY		
Ref.No.		
Advertisement Ref. No.		Date :

NO	TERMS AND CONDITIONS	**VENDOR'S OFFER** (PLEASE STATE)
1.0	Tenderer must be registered with the Ministry of Health.	
2.0	**TENDER FORM should be filled** completely including the **USER REQUIREMENT FORM** (if available). Submission of incomplete form **MAY** cause **DISQUALIFICATION OF TENDER.**	
3.0	Each tenderer is allowed to quote **ONE BRAND WITH ONE PRICE ONLY** for each item. Submission of more than one brand and price will cause **DISQUALIFICATION OF TENDER.**	
4.0	All consumables supplied throughout this tender shall have a minimum expiry date of twelve (12) months / on delivery. Should the consumbles be urgently needed, provision of consumbles with expiry date of less than twelve (12) months should be first agreed by the User before delivery is made.	
5.0	**Brochures / catalogues should be submitted / attached with tender document.**	
6.0	**SERVICE AND AFTER SALES SUPPORT**	
6.1	All consumables supplied throughout this tender shall have a minimum expiry date of **six (6) months on delivery**. Should the consumable be urgently needed, provision of a consumable with expiry date of less than six (6) months should be first agreed by the User of the particular laboratory before delivery is made.	
6.2	Letter of Undertaking (LOU) shall be produced upon each delivery of consumable with expiry date of less than **six (6)** months and vendor shall declare in the LOU that unused, unopened, expired kits will be replaced accordingly.	
6.3	User shall have the rights to return any items, and to be replaced at no extra cost, if found not meeting the acceptance criteria upon opening a pack such as, but not limited to, the following: 1. Tampered or damaged packaging 2. Evident of leakage or damaged products 3. Expired products that are evidently less than the requirement mentioned in para 4.1 calculated from delivery date 4. Leakage upon delivery	

6.4	Vendor shall submit samples of the offered items directly to the Users no later than 7 days after the Closing Date of this advertisement or as required by the Users.	
7.0	EXIT CLAUSE: The tender contract shall be automatically terminated even though tender has not yet expired and this shall be in effect due to, but not limited to, the following: **1.** When the testing is no longer required or relevant i.e. test is obsolete, to the laboratory or department. **2.** When the item(s) set out in this tender is/are no longer required by the laboratory or the Department. **3.** When the approved budget allocation for this tender contract has been used up before the tender contract expires whereby a renewal of tender shall be submitted by the user for an open advertisement subject to approval by the Mini Tender Board (*Lembaga Tawaran Kecil*).	
8.0	DELIVERY PERIOD: **Preferably 4 – 8 weeks and no later than 12 weeks after issue of Purchase Order**	(Yes / No) **(If No, please specify)**
9.0	PRICE VALIDITY: The quotation shall remain valid for **12 months** from the final date for the submission of the quotation and no supplier may withdraw his/her quotation within that period. The Government reserves the right to extend this period if deemed necessary provided that such extension to the quotation validity period shall have written consent of the supplier(s).	

We offer and undertake on your acceptance of our Tender to supply and deliver the above mentioned goods in accordance with your Invitation To Tender.

2. Our Tender is fully consistent with and does not contradict or derogate from anything in your Invitation To Tender. We have not qualified or changed any of the provisions of your Invitation To Tender.

3. We shall execute a formal agreement in the appropriate form set out in Section 4 – Contract of the Invitation to Tender together with such further terms and conditions, if any, agreed between the Government and us.

4. OUR OFFER IS VALID FOR **TWELVE (12) CALENDER** MONTHS FROM THE TENDER CLOSING DATE.

5. When requested by you, we shall extend the validity of this offer.

6. We further undertake to give you any further information which you may require.

Dated this day of 20 .

_____ Tenderer's official stamp:

Signature of authorised officer of Tenderer
Name:
Designation:

SCHEDULE 2 - INFORMATION SUMMARY

2.1 Tenderers shall provide in this Schedule the following information:

a. Management summary

b. Company profile (including Contractor and sub-contractor(s), if any)

c. Years of experience (as of the Tender Closing Date) of the Contractor and sub-contractor(s) in the:

- *Supply, Delivery Medical Consumables*

d. Other information which is considered relevant

SCHEDULE 3 – SUB-CONTRACTS

3.1 Tenderers shall complete Table 3.1 with information about all the companies involved in the provision of the services and items specified in this tender. This shall include details about the Contractor and each sub-contractor involved, as well as their respective responsibilities.

3.2 Tenderers shall also indicate in Table 3.1 any alliance relationship established with each sub-contractor. An alliance is defined as a formal and binding business relationship between the allied parties.

Table 3.1 - Responsibility Table

Company Name	Responsibility Description	Alliance Relationship between Contractor and Sub-contractor(s)		
		Alliance Exists? (Y/N)	Date Established	Alliance Description
Contractor				
		Not Applicable	Not Applicable	Not Applicable
Sub-contractor(s)				

SCHEDULE 4 – COMPANY'S BACKGROUND

4.1 Each of the companies involved in this tender, including Contractor and sub-contractor(s) (if any), shall provide information on the company's background, scope of operations, financial standing and certified copy of its Certificate of Incorporation or Certificate of Registration (as the case may be).

SCHEDULE 5 – REFERENCES

5.1 Tenderers shall submit a list of customers in Table 5.1 to whom the Contractor has provided similar services and items as specified in this tender in the recent 5 years as of the Tender Closing Date.

Table 5.1 - References of previous customers

Customer Name and Address	Customer Type (Govt or Quasi Govt)*	Contact Person	Title	Contact Number, Fax Number and E-mail Address

***Note: Tenderers shall indicate whether the customer is a Government or Quasi Government organisation. A Quasi Government is defined as an organisation which (1) is managed and controlled by the Government; or (2) has at least 50% shares being held by the Government. Please leave the column blank if the customer is neither a Government or Quasi Government organisation.**

5.2 The Ministry of Health shall treat all the information submitted under this schedule in strict confidence.

5.3 The Ministry of Health reserves the right to contact the references for tender assessment purposes.

SCHEDULE 6 - SUBMISSION OF SAMPLE

6.1 Tenderers shall submit the Submission of Sample form below in respect of the items specified in this tender.

6.2 Samples of the items to be submitted shall be:

a. identical in packing and manufacture to the items to be offered by the Tenderer; and

b. marked with the corresponding item number of the tender.

SUBMISSION OF SAMPLE FORM

To:

TENDER REFERENCE NO: KK/264/2026/UPP(TC)

INVITATION TO TENDER
TO SUPPLY AND DELIVERY OF MEDICAL CONSUMABLES FOR HAEMATOLOGY UNIT, RAJA ISTERI PENGIRAN ANAK SALEHA (RIPAS) HOSPITAL, MINISTRY OF HEALTH FOR A PERIOD OF THREE (3) YEARS USAGE

SUBMISSION OF SAMPLE FORM OF (*NAME OF TENDERER*)

_____

ITEM NO.	DESCRIPITON	SAMPLE SUBMITTED (indicate with ✔)	SAMPLE NOT SUBMITTED (indicate with ✕)	OFFERED/ NOT OFFERED (indicate as appropriate)
1.	Trima Accel® LRS® Platelet, Plasma Set			
2.	Anticoagulant Citrate Dextrose Solution USP(ACP) Solution A, 750ml			
3.	Spectra Optia™ Exchange Set			
4.	Spectra Optia™ IDL Set			

We understand as stated in the Instructions To Tenderers that Tenders without samples shall not be considered.

_____ Tenderer's official stamp:
(*signature of authorized officer of Tenderer*)
Name:
Designation:
Date:

_____

FOR OFFICE USE

Date of receipt : _____

Receiving Officer : _____